



## Division of Criminal Justice Services

---

### CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 129906, performed on September 19, 2019 by Office of Public Safety employee Joseph Iavarone, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under  
ESRA by Joshua Vinehout  
on 9/25/2019 8:49 AM

---

**Joshua Vinehout**  
Supervisor of Public Safety Programs  
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Liverpool Village Police Department  
Manufacturer: NAT'L PATENT ANALYTICAL SYS      Model: DATAMASTER DMT  
Date: September 19, 2019      Serial Number: 129906

Reference Solution Lot Number: 18371  
Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.100 %	33.99 degrees centigrade
2	0.099 %	33.99 degrees centigrade
3	0.100 %	33.99 degrees centigrade
4	0.100 %	33.99 degrees centigrade

Reference Solution Lot Number: 19101  
Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.081 %	33.99 degrees centigrade
2	0.081 %	33.99 degrees centigrade
3	0.081 %	33.99 degrees centigrade
4	0.081 %	33.99 degrees centigrade

Reference Solution Lot Number: 19210  
Reference Solution Standard: 0.180 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.184 %	33.99 degrees centigrade
2	0.182 %	33.99 degrees centigrade
3	0.183 %	33.99 degrees centigrade
4	0.183 %	33.99 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 129906, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.


Performed by:



Digitally signed under  
ESRA by Joseph Iavarone  
on 9/19/2019 3:31 PM

**Joseph Iavarone**  
Highway Safety Equipment Technician  
Breath Analysis Operator Permit # 44106

Reviewed by:



Digitally signed under  
ESRA by Frank Sgarlata  
on 9/23/2019 8:23 AM

**Frank Sgarlata**  
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Liverpool Village Police Department  
Manufacturer: NAT'L PATENT ANALYTICAL SYS  
Model: DATAMASTER DMT  
Serial Number: 129906  
Date: September 19, 2019

1. General Appearance and condition of this device -----: OK

2. Check controls and indicators -----: OK

3. Check breath flow system -----: OK

4. Check printer operation -----: OK

5. System cycles through all modes -----: OK

6. Record any repairs made to this instrument and list services provided: **INSPECTED INSTRUMENT, CHECKED SETUP VOLTAGES PRINT AND REMOTE OPERATION. RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION.**

Performed by:



Digitally signed under  
ESRA by Joseph Iavarone  
on 9/19/2019 3:31 PM

---

**Joseph Iavarone**  
Highway Safety Equipment Technician  
Breath Analysis Operator Permit # 44106

Reviewed by:



Digitally signed under  
ESRA by Frank Sgarlata  
on 9/23/2019 8:23 AM

---

**Frank Sgarlata**  
Breath Analysis Technical Supervisor



**Division of Criminal  
Justice Services**

---

**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS  
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 129906, performed on March 12, 2019 by Office of Public Safety employee Joseph Iavarone, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance/calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under ESRA  
by Joshua Vinehout  
on 3/18/2019 9:57 AM

---

**Joshua Vinehout  
Supervisor of Public Safety Programs  
Office of Public Safety**

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Liverpool Village Police Department  
Manufacturer: NAT'L PATENT ANALYTICAL SYS      Model: DATAMASTER DMT  
Serial Number: 129906  
Date: March 12, 2019

Reference Solution Lot Number: 18360  
Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.100 %	34.00 degrees centigrade
2	0.100 %	34.00 degrees centigrade
3	0.101 %	34.00 degrees centigrade
4	0.101 %	34.00 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 129906, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.


Performed by:



Digitally signed under  
ESRA by Joseph Iavarone  
on 3/12/2019 3:34 PM

**Joseph Iavarone**  
Highway Safety Equipment Technician  
Breath Analysis Operator Permit # 44106

Reviewed by:



Digitally signed under  
ESRA by Frank Sgarlata  
on 3/14/2019 12:44 PM

**Frank Sgarlata**  
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Name of Submitting Agency: Liverpool Village Police Department**  
**Manufacturer: NAT'L PATENT ANALYTICAL SYS**  
**Model: DATAMASTER DMT**  
**Serial Number: 129906**  
**Date: March 12, 2019**

1. General Appearance and condition of this device -----: N/A

2. Check controls and indicators -----: OK

3. Check breath flow system -----: OK

4. Check printer operation -----: N/A

5. System cycles through all modes -----: OK

6. Record any repairs made to this instrument and list services provided: **CHECKED REMOTE OPERATION, RAN DIAGNOSTIC AND TECHNICIAN TESTS, VERIFIED CALIBRATION.**

Performed by:

*Joseph Iavarone*

Digitally signed under  
ESRA by Joseph Iavarone  
on 3/12/2019 3:34 PM

**Joseph Iavarone**  
**Highway Safety Equipment Technician**  
**Breath Analysis Operator Permit # 44106**

Reviewed by:

*Frank Sgarlata*

Digitally signed under  
ESRA by Frank Sgarlata  
on 3/14/2019 12:44 PM

**Frank Sgarlata**  
**Breath Analysis Technical Supervisor**



## Division of Criminal Justice Services

### CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 129906, performed on September 21, 2018 by Office of Public Safety employee Michael J. Canary, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under  
ESRA by Joshua Vinehout  
on 9/26/2018 8:09 AM

---

Joshua Vinehout  
Supervisor of Public Safety Programs  
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Liverpool Village Police Department  
Manufacturer: NAT'L PATENT ANALYTICAL SYS      Model: DATAMASTER DMT  
Date: September 21, 2018      Serial Number: 129906

Reference Solution Lot Number: 18010  
Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.099 %	34.03 degrees centigrade
2	0.098 %	34.03 degrees centigrade
3	0.099 %	34.02 degrees centigrade
4	0.099 %	34.03 degrees centigrade

Reference Solution Lot Number: 18161  
Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.083 %	34.02 degrees centigrade
2	0.083 %	34.03 degrees centigrade
3	0.083 %	34.02 degrees centigrade
4	0.082 %	34.02 degrees centigrade

Reference Solution Lot Number: 17330  
Reference Solution Standard: 0.180 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.181 %	34.02 degrees centigrade
2	0.182 %	34.02 degrees centigrade
3	0.181 %	34.03 degrees centigrade
4	0.181 %	34.02 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 129906, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:



Digitally signed under  
ESRA by Michael J. Canary  
on 9/21/2018 12:47 PM

**Michael J. Canary**  
Highway Safety Equipment Technician  
Breath Analysis Operator Permit # 41924

Reviewed by:




Digitally signed under  
ESRA by Frank Sgarlata  
on 9/25/2018 8:42 AM

**Frank Sgarlata**  
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Liverpool Village Police Department  
Manufacturer: NAT'L PATENT ANALYTICAL SYS  
Model: DATAMASTER DMT  
Serial Number: 129906  
Date: September 21, 2018

- 1. General Appearance and condition of this device -----: OK
- 2. Check controls and indicators -----: OK
- 3. Check breath flow system -----: OK
- 4. Check printer operation -----: OK
- 5. System cycles through all modes -----: OK



6. Record any repairs made to this instrument and list services provided: INIT. INSPECTION. PRINT AND REMOTE OPERATION. CHECK SET-UP VOLTS. PERFORM CAL/ADJ PROCESS. PERFORM SUBJECT TESTS, DIAG. TESTS AND TECH. TESTS. PIC (2.08) / SW - 2.02-5- VERIFIED CALIBRATION.

Performed by:

Handwritten signature of Michael J. Canary.

Digitally signed under  
ESRA by Michael J. Canary  
on 9/21/2018 12:47 PM

---

Michael J. Canary  
Highway Safety Equipment Technician  
Breath Analysis Operator Permit # 41924

Reviewed by:

Handwritten signature of Frank Sgarlata.

Digitally signed under  
ESRA by Frank Sgarlata  
on 9/25/2018 8:42 AM

---

Frank Sgarlata  
Breath Analysis Technical Supervisor



Division of Criminal  
Justice Services

---

**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS  
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 129906, performed on March 20, 2018 by Office of Public Safety employee Michael J. Hess, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance/calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under ESRA  
by Joshua Vinehout  
on 3/26/2018 1:02 PM

---

**Joshua Vinehout**  
Supervisor of Public Safety Programs  
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Liverpool Village Police Department  
Manufacturer: NAT'L PATENT ANALYTICAL SYS      Model: DATAMASTER DMT  
Serial Number: 129906  
Date: March 20, 2018

Reference Solution Lot Number: 17390  
Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.097 %	34.0 degrees centigrade
2	0.098 %	34.0 degrees centigrade
3	0.098 %	34.0 degrees centigrade
4	0.098 %	34.0 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 129906, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:



Digitally signed under  
ESRA by Michael J. Hess  
on 3/20/2018 1:25 PM

---

Michael J. Hess  
Highway Safety Equipment Technician  
Breath Analysis Operator Permit # 16080

Reviewed by:



Digitally signed under  
ESRA by Frank Sgarlata  
on 3/26/2018 8:25 AM

---

Frank Sgarlata  
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Name of Submitting Agency:** Liverpool Village Police Department  
**Manufacturer:** NAT'L PATENT ANALYTICAL SYS  
**Model:** DATAMASTER DMT  
**Serial Number:** 129906  
**Date:** March 20, 2018

1. General Appearance and condition of this device -----: N/A

2. Check controls and indicators -----: OK

3. Check breath flow system -----: OK

4. Check printer operation -----: OK

5. System cycles through all modes -----: OK

6. Record any repairs made to this instrument and list services provided: RAN COMMUNICATION, DIAGNOSTIC AND SUPERVISOR TESTS, VERIFIED CALIBRATION



Performed by:

Digitally signed under  
ESRA by Michael J. Hess  
on 3/20/2018 1:25 PM

**Michael J. Hess**  
**Highway Safety Equipment Technician**  
**Breath Analysis Operator Permit # 16080**

Reviewed by:

Digitally signed under  
ESRA by Frank Sgarlata  
on 3/26/2018 8:25 AM

**Frank Sgarlata**  
**Breath Analysis Technical Supervisor**



## Division of Criminal Justice Services

### **CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 129906, performed on September 25, 2017 by Office of Public Safety employee Martin Schilling, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under  
ESRA by Joshua Vinehout  
on 9/27/2017 2:17 PM

---

**Joshua Vinehout**  
Supervisor of Public Safety Programs  
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Liverpool Village Police Department  
 Manufacturer: NAT'L PATENT ANALYTICAL SYS      Model: DATAMASTER DMT  
 Date: September 25, 2017      Serial Number: 129906

Reference Solution Lot Number: 17061  
 Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.101 %	34.01 degrees centigrade
2	0.100 %	34.00 degrees centigrade
3	0.100 %	34.01 degrees centigrade
4	0.100 %	34.01 degrees centigrade

Reference Solution Lot Number: 17070  
 Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.078 %	34.00 degrees centigrade
2	0.078 %	34.01 degrees centigrade
3	0.078 %	34.00 degrees centigrade
4	0.078 %	34.01 degrees centigrade

Reference Solution Lot Number: 16370  
 Reference Solution Standard: 0.180 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.177 %	34.00 degrees centigrade
2	0.177 %	34.01 degrees centigrade
3	0.178 %	34.01 degrees centigrade
4	0.177 %	34.00 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 129906, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.


I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:

Reviewed by:



Digitally signed under  
 ESRA by Martin Schilling  
 on 9/25/2017 1:04 PM



Digitally signed under  
 ESRA by Frank Sgarlata  
 on 9/27/2017 9:21 AM

**Martin Schilling**  
 Highway Safety Equipment Technician  
 Breath Analysis Operator Permit # 45926

**Frank Sgarlata**  
 Highway Safety Equipment Technical Supervisor  
 Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Name of Submitting Agency: Liverpool Village Police Department**  
**Manufacturer: NAT'L PATENT ANALYTICAL SYS**  
**Model: DATAMASTER DMT**  
**Serial Number: 129906**  
**Date: September 25, 2017**

- 1. General Appearance and condition of this device -----: OK
- 2. Check controls and indicators -----: OK
- 3. Check breath flow system -----: OK
- 4. Check printer operation -----: OK
- 5. System cycles through all modes -----: OK

**6. Record any repairs made to this instrument and list services provided: INSPECTED INSTRUMENT, CHECKED SETUP VOLTAGES PRINT AND REMOTE OPERATION. PERFORMED CALIBRATION/ADJUSTMENT PROCEDURE, RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION.**

Performed by:



Digitally signed under  
ESRA by Martin Schilling  
on 9/25/2017 1:04 PM

---

**Martin Schilling**  
**Highway Safety Equipment Technician**  
**Breath Analysis Operator Permit # 45926**

Reviewed by:



Digitally signed under  
ESRA by Frank Sgarlata  
on 9/27/2017 9:21 AM

---

**Frank Sgarlata**  
**Highway Safety Equipment Technical**  
**Supervisor**  
**Breath Analysis Technical Supervisor**



**Division of Criminal  
Justice Services**

---

**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS  
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Matthew J. Slater, Associate Training Technician (Police), Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 129906, performed on February 22, 2017 by Office of Public Safety employee Michael J. Hess, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance/calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Matthew J. Slater".

Digitally signed under ESRA  
by Matthew J. Slater  
on 2/24/2017 2:26 PM

---

**Matthew J. Slater**  
**Associate Training Technician (Police)**  
**Office of Public Safety**

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Liverpool Village Police Department  
Manufacturer: NAT'L PATENT ANALYTICAL SYS      Model: DATAMASTER DMT  
Serial Number: 129906  
Date: February 22, 2017

Reference Solution Lot Number: 16380  
Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.099 %	34.0 degrees centigrade
2	0.099 %	34.0 degrees centigrade
3	0.100 %	34.0 degrees centigrade
4	0.101 %	34.0 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 129906, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:



Digitally signed under  
ESRA by Michael J. Hess  
on 2/23/2017 7:46 AM

**Michael J. Hess**  
Highway Safety Equipment Technician  
Breath Analysis Operator Permit # 16080

Reviewed by:



Digitally signed under  
ESRA by Lisa A. Lee  
on 2/24/2017 8:44 AM

**Lisa A. Lee**  
Highway Safety Equipment Technical Supervisor  
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Name of Submitting Agency:** Liverpool Village Police Department  
**Manufacturer:** NAT'L PATENT ANALYTICAL SYS  
**Model:** DATAMASTER DMT  
**Serial Number:** 129906  
**Date:** February 22, 2017

1. General Appearance and condition of this device -----: N/A

2. Check controls and indicators -----: OK

3. Check breath flow system -----: OK

4. Check printer operation -----: OK

5. System cycles through all modes -----: OK

6. Record any repairs made to this instrument and list services provided: RAN COMMUNICATION, DIAGNOSTIC AND SUPERVISOR TESTS, VERIFIED CALIBRATION

Performed by:

 Digitally signed under  
ESRA by Michael J. Hess  
on 2/23/2017 7:46 AM

**Michael J. Hess**  
**Highway Safety Equipment Technician**  
**Breath Analysis Operator Permit # 16080**

Reviewed by:

 Digitally signed under  
ESRA by Lisa A. Lee  
on 2/24/2017 8:44 AM

**Lisa A. Lee**  
**Highway Safety Equipment Technical Supervisor**  
**Breath Analysis Technical Supervisor**